

Declaration of Conformity

This European Declaration of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER			
Name of Company		Address	SRN
Mirion Technologies (Capintec), Inc.		7 Vreeland Road, Florham Park, NJ 09732 USA	US-MF-000000770
AUTHORIZED REPRESENTATIVE			
Name of Company	Address	SRN	Telephone/fax/email
Atlantico Systems, Ltd.	34 Oldfield, Galway, Ireland	IE-AR-000000208	+35391443609/ info@AtlanticoSystems.com
PRODUCT IDENTIFICATION			
Product Name		Product Code / Catalog Number	Basic UDI-DI/EMDN
CRC 55tR Dose Calibrator		5130-3234	08599420061004Q/ Z11029005
CRC 55tPET Dose Calibrator		5130-3235	
CRC 55tW Dose Calibrator		5130-2216	
Intended Purpose			Photo
The CRC Dose Calibrators are intended to measure radiopharmaceuticals or brachytherapy sources prior to administration to patients.			NA
RISK CLASS FOR MEDICAL DEVICES			
Device Classification	Standards		
Class: Im Rule: 13, Annex VIII Conformity Assessment per Annex IX	IEC 60601-1-2 Ed. 4.0, (2014): Medical Electrical Equipment – Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests ANSI/AAMI ES60601-1:2005/A1:2012 Issued: 2012/08/20 Medical electrical equipment – Part 1: General requirements for basic safety and essential performance CAN/CSA-C22.2 No. 60601-1:14: Third Edition Issued: 2014/03/01 – Medical electrical equipment – Part 1: General requirements for basic safety and essential performance		



	IEC 60601-1-6 Issued: 2013/10/29 Ed: 3.1 Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral Standard: Usability IEC 62366:2014, Edition 1.1 Issued 2014/01/28 – Medical devices – Application of usability engineering to medical devices IEC 62304:2006 + A1:2015 issued Medical device software – Software life cycle processes		
NOTIFIED BODY			
Name of Company	ID Number	Address	Certificate Reference(s)
Intertek Medical Notified Body AB	2862	Torshamnsgatan 43 Box 1103, SE-164 22 Kista, Sweden	286201185244

Autograph declares that the above-mentioned products meet the provision of the following EU legislation:

- MDR (EU) 2017/745 (LVFS 2003:11 as amended by LVFS 2009:07)

COMPANY REPRESENTATIVE:

Mary Anne Yusko

TITLE: Director, Regulatory Affairs



SIGNATURE:

DATE September 9, 2024

PLACE: Mirion Technologies (Capintec), Inc.

7 Vreeland Road, Florham Park, NJ 07932 USA

