



CAPINTEC, INC.

DECLARATION OF CONFORMITY THE EC DIRECTIVE AND CE MARKING

We, Mirion Technologies (Capintec), Inc.
7 Vreeland Road
Florham Park, NJ 07932 USA
Phone (800) ASK-4CRC
Fax (201) 825-1336

declare under our sole responsibility that the product Caprac®-t Well Counter (including use with Epson Roll and Epson Slip printers) to which this declaration relates is in conformity with the Low Voltage Directive 2014/95/EC and Electromagnetic Compatibility 2014/30/EU

This product is in conformity with the following standards or other normative documents following:

- IEC 60601-1-2 Ed. 4.0, (2014): Medical Electrical Equipment – Part 1-2: General requirements for basic safety and essential performance - Collateral Standard: Electromagnetic disturbances - Requirements and tests
- IEC 60601-1:2005, AMD1:2012 Issued: 2012/08/20 Medical electrical equipment – Part 1: General requirements for basic safety and essential performance
- CAN/CSA-C22.2 No. 60601-1:14: Third Edition Issued: 2014/03/01 – Medical electrical equipment – Part 1: General requirements for basic safety and essential performance
- IEC 60601-1-6:2010, AMD1:2013 Issued: 2013/10/29 Ed: 3.1 Medical electrical equipment – Part 1-6: General requirements for basic safety and essential performance – Collateral Standard: Usability
- IEC 62366-1:2015, Edition 1.1 Issued 2014/01/28 – Medical devices – Application of usability engineering to medical devices
- IEC 62304:2006 + A1:2015 Ed.1.0 Medical device software – Software life cycle processes

CE Mark as shown denotes conformance to the above statement.



The Technical Construction File is maintained at:

Mirion Technologies (Capintec), Inc. Phone: (800) ASK-4CRC
7 Vreeland Road Fax: (201) 825-1336
Florham Park, NJ 07932 USA

Date of issue: 4/11/2021

Place of issue:

Mirion Technologies (Capintec), Inc.
7 Vreeland Road
Florham Park, NJ 07932 USA

Mary Anne Yusko
Vice President, Product Development & Regulatory Affairs